

**Angle rotor, 12-place**


**Hematocrit rotor, 24-place**



∠ 35°


n = 12,000 min<sup>-1</sup>  
max. RCF 16,582



**Cat. No. 1615**



n = 15,000 min<sup>-1</sup>  
max. RCF 21,382

**Cat. No. 1650**

capacity in ml	4	5	6	15	1.1–1.4	2.6–3.4	2.7–3	4.5–5	4.9	7.5–10	10	1.6–5	4–7	8–10	15	
∅ x L in mm	10x88	12x75	12x82	17x100	8x66	13x65	11x66	11x92	13x90	15/16x92	15x102	13x75	13x100	16x100/125	17x120	
<b>Cat. No.</b>	-	<b>0553<sup>2)</sup></b>	<b>0501<sup>2)</sup></b>	<b>0518<sup>2)</sup></b>	<b>blood collection/urine tubes</b>										<b>0509</b>	
 <b>rotor</b> Cat. No. <b>1615</b>																
<b>Cat. No.</b>	<b>6305</b>	<b>1054-A</b>	-	<b>1054-A</b>	-						<b>1054-A</b>	<b>1058</b>	-	<b>1647<sup>29)</sup></b>		
boring ∅ x L in mm	11.5x67.5	13.5x60	17.7x88	13.5x60	17.7x88						13.5x60	13.5x79	17.7x88	17x104		
tubes per rotor					12						12	12	12	6	6	
max. RCF <sup>2)</sup>	14,006	13,201	16,582	13,201	16,582						13,201	16,582	16,582	15,455		
radius in mm	87	82	103	82	103						82	103	103	96		
run-up in sec	40															
run-down in sec, braked	40															
temperature in °C <sup>1)</sup>	-2															

capillaries, heparinised	basic	self-sealing and mylar-coated
<b>Cat. No.</b>	<b>2074</b>	<b>1071</b>
lid as evaluation disk incl.		
 <b>rotor</b> Cat. No. <b>1650</b>	sealing putty	
<b>Cat. No.</b>	<b>2077</b>	-
boring ∅ x L in mm	-	
capillaries per rotor	24	
max. RCF <sup>2)</sup>	21,382	
radius in mm	85	
run-up in sec	13	
run-down in sec, braked	9	
temperature in °C <sup>1)</sup>	-1	